


GRAPH LEGEND



TESTING THE MUSCLE

1. Dry muscles
2. Relaxed state during test
3. Sensor placement refer to electromyography book illustrations
4. Place electrodes proximal and distal of sensor on selected muscle
5. Start at 20mA and go up gradually until the contraction time and displacement have no or minimal change
6. Make sure that the measured subject (joints) is always in standardized position

MUSCLE SCENARIOS

